

# How Industry Utilizes GPS for Traceable Frequency Measurements and Calibrations


Jeff Gust

Chief Corporate Metrologist


Fluke Corp.

# Where Frequency Measurements Are Needed

- Testing Laboratories (Biomedical, Chemical)
- Calibration Laboratories
- Legal Metrology
- Communications
- Industrial Process Control
- Electrical
- Police Radar
- Power/Energy



- Metrological Traceability to the SI (International System of Units)




# Metrological Traceability Defined

FLUKE®

Calibration

- International Vocabulary of Metrology (VIM)

Property of a measurement result whereby the result can be related to a reference through a documented unbroken chain of calibrations, each contributing to the measurement uncertainty



- Circular T report available at BIPM Website
- <http://www.bipm.org/en/bipm-services/timescales/time-ftp/cirt.html#nohref>
- NIST uncertainty  $\approx 1 \times 10^{-14}$

CIRCULAR T 344  
2016 SEPTEMBER 07, 11h UTC

ISSN

BUREAU INTERNATIONAL DES POIDS ET MESURES  
ORGANISATION INTERGOUVERNEMENTALE DE LA CONVENTION DU METRE  
PAVILLON DE BRETEUIL F-92312 SEVRES CEDEX TEL. +33 1 45 07 70 70 FAX. +33 1 45 34 20 21 tai@bipm.org

The contents of the sections of BIPM *Circular T* are fully described in the document " [Explanatory supplement to BIPM Circular T](ftp://ftp2.bipm.org/pub/tai/publication/notes/explanatory_supplement_v0.1.pdf) " available at [ftp://ftp2.bipm.org/pub/tai/publication/notes/explanatory\\_supplement\\_v0.1.pdf](ftp://ftp2.bipm.org/pub/tai/publication/notes/explanatory_supplement_v0.1.pdf)

**1** - Difference between UTC and its local realizations UTC(k) and corresponding uncertainties. From 2015 July 1, 0h UTC, to 2017 January 1, 0h UTC,  $TAI-UTC = 36$  s. From 2017 January 1, 0h UTC,  $TAI-UTC = 37$  s.

Date 2016 0h UTC	MJD	JUL 30	AUG 4	AUG 9	AUG 14	AUG 19	AUG 24	AUG 29	Uncertainty/ns	" <sub>A</sub>	" <sub>B</sub>	"
		57599	57604	57609	57614	57619	57624	57629				

- GPS Data Archive maintained by NIST
- <https://www.nist.gov/pml/time-and-frequency-division/services/gps-data-archive>
- NIST Uncertainty  $\approx 1 \times 10^{-14}$

### GPS monitoring data for the 30 day period ending 2016-08-31 (as received at NIST in Boulder, Colorado)

[Archive Home](#)

[1 Day Averages](#)

[1 Hour Averages](#)

[10 Minute Averages](#)

[Next Date](#)

[Last Date](#)

GPS - UTC(NIST)  
(one-hour averages using all satellites in view)

Hours	Mean Time Offset (ns)	Range (ns)	Frequency Offset	Confidence (r)
720	-8.20	19.98	$<1.0 \times 10^{-15}$	+0.17

# GPS Data Archive

GPS PRN - UTC(NIST)  
(data from individual GPS satellites)


GPS PRN	Minutes (In-View)	Mean Time Offset	Range (ns)	Time Deviation	Frequency Offset
1	12180	-6.31	33.95	2.38	<1.0 x 10 <sup>-15</sup>
2	10920	-12.62	36.00	1.56	<1.0 x 10 <sup>-15</sup>
3	11060	-5.62	47.05	2.77	<1.0 x 10 <sup>-15</sup>
4	---	---	---	---	---
5	9340	-8.49	39.45	1.94	<1.0 x 10 <sup>-15</sup>
6	9860	-7.22	35.35	1.90	<1.0 x 10 <sup>-15</sup>
7	11420	-6.54	33.60	1.78	+1.0 x 10 <sup>-15</sup>
8	11040	-7.26	32.80	2.52	+1.3 x 10 <sup>-15</sup>
9	9650	-4.74	38.95	1.67	+2.6 x 10 <sup>-15</sup>
10	11110	-6.99	141.60	3.23	+1.7 x 10 <sup>-15</sup>
11	11140	-9.80	103.10	3.49	<1.0 x 10 <sup>-15</sup>
12	9940	-8.28	40.30	1.67	<1.0 x 10 <sup>-15</sup>
13	12730	-7.46	39.85	2.46	<1.0 x 10 <sup>-15</sup>
14	11230	-9.63	34.45	1.79	<1.0 x 10 <sup>-15</sup>
15	10020	-8.54	58.30	2.70	<1.0 x 10 <sup>-15</sup>
16	10150	-8.42	34.05	2.41	+1.6 x 10 <sup>-15</sup>
17	8540	-7.72	28.55	1.62	<1.0 x 10 <sup>-15</sup>
18	12000	-11.08	140.60	2.19	+1.3 x 10 <sup>-15</sup>
19	10430	-12.32	33.60	1.91	<1.0 x 10 <sup>-15</sup>
20	9050	-12.24	48.35	1.81	<1.0 x 10 <sup>-15</sup>
21	11850	-12.23	32.30	1.61	<1.0 x 10 <sup>-15</sup>
22	8000	-9.95	23.55	1.36	+1.2 x 10 <sup>-15</sup>
23	10590	-7.61	38.00	1.69	+1.8 x 10 <sup>-15</sup>
24	13070	-6.51	40.40	2.32	<1.0 x 10 <sup>-15</sup>
25	10310	-6.69	28.25	1.92	<1.0 x 10 <sup>-15</sup>
26	11110	-5.35	47.15	2.22	+1.8 x 10 <sup>-15</sup>
27	12150	-6.14	45.15	2.36	+1.7 x 10 <sup>-15</sup>
28	12350	-10.42	37.50	1.69	<1.0 x 10 <sup>-15</sup>
29	9750	-8.99	34.15	1.67	<1.0 x 10 <sup>-15</sup>
30	9790	-7.43	33.35	1.76	+1.4 x 10 <sup>-15</sup>
31	9720	-5.21	33.55	2.30	+1.2 x 10 <sup>-15</sup>
32	8780	-6.23	37.30	1.54	+1.3 x 10 <sup>-15</sup>

- GPS Disciplined Oscillator
- Fluke 910R
- Freq. offset (24 hour mean)  $< 1 \times 10^{-12}$
- Allan Deviation  $< 1 \times 10^{-12}$  (t = 100 s)





- The only GPSDSO that provides cal data of the oscillator from GPS
- Data is stored and available to the user



# Verification of Performance

- Fluke has a reference 910R that is continuously compared using the NIST FMAS
- Uncertainty  $\approx 3 \times 10^{-13}$
- Frequency offset is measured by FMAS and data is compared to the 910R internal data for agreement to  $1 \times 10^{-12}$
- 910R owners can send in their units for calibrations
- Certificates issued are under laboratory scope of accreditation to ISO/IEC 17025



# GPSDSO to Counters and Sources

FLUKE®


Calibration

- Time Base oscillators for Frequency Counters and Signal Generators calibrated via frequency comparison to GPSDSO
- Uncertainty  $\approx 1 \times 10^{-11}$



# GPSDSO as House Frequency Std

- Josephson Voltage Standard, primary standard for DC volt
- $K_{J-90}$  is assigned as 483 597.9 GHz/V
- Key contributor to uncertainty is frequency standard that drives Gunn diode at  $\approx 75$  GHz
- Uncertainty r



# Down the Traceability Chain

- Frequency Counter calibrates lower accuracy Signal Generators
- Signal Generators calibrate lower accuracy Frequency measurement devices
- Timometer (stop watch calibrator)
- Specified accuracy 0.05 s/day ( $\approx 5.8 \times 10^{-7}$ )



# Conclusion

- GPSDO's are used in nearly every frequency calibration laboratory around the world
- Industrial measurement traceability depends on GPS